



Item number 3895



Excellent

Examination gloves, ABENA Excellent, XL, nature, latex, powder free, double chlorinated

- ✓ Smooth surface
- ✓ Elastic
- ✓ Good fit
- ✓ Good sensitivity



Product Description

Elastic examination glove in natural latex material with an extra low latex protein content. Good fit, providing the user high sensitivity.



LOW
GRIP

AQL
1.5

EN EN
420 455

Specifications

Base name	Examination gloves
Brand	ABENA
Sub-brand	Excellent
Size	XL
Sub-color	Nature
Properties	Full textured, low grip, rolled cuff
Features	Powder free, double chlorinated
Single or multiple use	Single use
Material	Latex
Technical ingredients	Latex, Sulfur, ZDEC, ZDBC, Zinc oxide, Potassium Hydroxide, Titanium Dioxide, Chlorine
Length/depth	240 mm
Width	115 mm
Thickness	0,12 +/- 0,02 mm
Certifications	CE, Food contact materials, MD
CE Category (Personal Protective Equipment)	CAT I
CE Class (Medical Devices)	Class I
Product or test standards	EN 455, AQL 1.5, EN 420
Directives, regulations and acts	MDR (EU) 2017/745, 1935/2004/EU, BEK 681 25.05.2020, 2023/2006/EC
Safety Instructions and Warnings	NB: if users or patients are hypersensitive to the material, wear vinyl or nitrile gloves. Contains latex proteins.
Because of the products composition it has an expiry date from the production date on:	3 years
Storage Instructions	Store dry, clean and at room temperature.
Product Disposal Instructions	Dispose with household waste.

Instructions for use/application

Examine the gloves for flaws and defects before use.

Packaging data

Unit	Contains	Length	Width	Height	EAN
cII	1000 pcs	44 cm	25.5 cm	25 cm	5703538115837
pck	100 pcs	24 cm	12 cm	8.5 cm	5703538115813
pair	2 pcs	0 cm	0 cm	0 cm	
pcs	1 pcs				



Regulation (EU) 2017/745 is a regulation of the European Union on the clinical investigation and sale of medical devices for human use. It repeals Directive 93/42/EEC, which concerns medical devices, and Directive 90/385/EEC, which concerns active implantable medical devices, on 26 May 2021.



The CE mark guarantees that a product is safe to use and complies with all safety precautions. CE stands for Conformité Européenne (European Conformity) and is mainly found on electronic equipment, safety equipment, construction products, and medical devices.



The glass-fork-symbol guarantees that products live up to the demands stated in the European legislation for food contact materials. The symbol is not mandatory but advised on products used for food contact.



The product contains latex.



The glove has a fully textured surface.



The glove is 240 mm long.



The glove has a low grip.



The standard specifies the requirements for protective gloves in relation to general requirements and test methods.



EN 455 consists of four standards that gloves must be tested against in order to be a medical glove. This standard specifies the requirements for disposable medical gloves.

